

EU Declaration of Conformity

Manufacturer:

Nitzbon AG

SRN: DE-MF-000009971

The manufacturer declares under its sole responsibility that the medical device:

Product name:

WP 3 wall system, Item no. 1622100

Intended purpose:

Occupational therapy, work simulation training, matching job profiles with skill profiles, and conducting occupational therapy and labour market-related performance analyses.


Basic UDI-DI: 4260101481622100UU

is in conformity with Regulation (EU) 2017/745 on medical devices.

The device is classified as Class I in accordance with Annex VIII, Rule 1 of Regulation (EU) 2017/745.

This declaration is valid until 31 December 2027 or until significant changes require revision of this declaration.

Hamburg, 1 January 2026
Nitzbon AG
Ergotechnik • Rehattechnik
Osterrade 14
21031 Hamburg
tel 0049 (0)40 739223-0
fax 0049 (0)40 73922399
email info@nitzbon.de



The logo consists of the word "NITZBON" in a bold, blue, sans-serif font, enclosed within a blue rectangular border that has a slight curve at the bottom. Below the logo is a handwritten signature in blue ink, which appears to be "Jens Nitzbon".

Jens Nitzbon

Annex to the Declaration of Conformity

Supplement for legacy products pursuant to Regulation (EU) 2017/745 (MDR, Art. 120)

1. Scope

This annex supplements the existing Declaration of Conformity dated January 1, 2026, for the following products that were placed on the market in accordance with Directive 93/42/EEC (MDD) and are subject to the transitional provisions of the MDR.

2. Affected products

Product Name	UDI-DI	Basis UDI-DI	Valid until
WP 3 wall system	4260101482013	4260101481622100UU	12/31/2027

Note: The product is a Class I device and has complied with the UDI requirements of the MDR since July 2021.

3. Legal Basis

- The products were placed on the market in accordance with MDD 93/42/EEC.
 - They are subject to the transitional provisions of the MDR 2017/745 (Art. 120) and may continue to be used until the specified date.
 - The UDI-DI and Basic UDI-DI comply with the requirements of the MDR (Annex VI).
-

4. Declaration

The manufacturer confirms that the above-mentioned products comply with the transitional provisions of the MDR and that the UDI requirements as well as post-market surveillance, vigilance, and market surveillance for legacy devices are met.

Hamburg, April 8, 2026

Jens Nitzbon PRRC

