



## DECLARATION OF CONFORMITY

Neuroquip Ltd  
St John's Innovation Centre  
Cowley Road  
Cambridge  
CB4 0WS  
United Kingdom

We hereby declare that the following device has been classified according to the classification rules and conforms to the essential principles for safety and performance specified in the Medical Devices Regulations EU 2017/745.

### **Manufacturing Site**

The Workshop  
Violet House  
Chapel Lane  
Elm  
Wisbech  
Cambridgeshire  
PE14 0DJ

### **Medical Device**

Neuroquip 'Action Research Arm Test Kit' – Model# ARAT001 / UDI: 5061008370002

### **Risk Classification**

Class 1 (Non-Sterile) Medical Device as classified in MDR Annex VIII, Chapter 3, Rule 4.1

### **Conformity Assessment**

Conformity assessment procedure: Class 1 Annex I and III of MDR (EU) 2017/745

This declaration of conformity is valid from 1<sup>st</sup> June 2018

*Andrew D Teager*

Andrew D Teager (Managing Director)

**Authorised Signatory**

Dated: 1<sup>st</sup> June 2018